TEKTRONIX®

465
OSCILLOSCOPE
AND
DM43/DM40
DIGITAL MULTIMETERS
OPERATORS

INSTRUCTION MANUAL

Tektronix, Inc.
P.O. Box 500
Beaverton, Oregon 97077

070-1738-01

Serial Number _____ First Printing OCT 1974

WARRANTY

All TEKTRONIX instruments are warranted against defective materials and workmanship for one year. Any questions with respect to the warranty should be taken up with your TEKTRONIX Field Engineer or representative.

All requests for repairs and replacement parts should be directed to the TEKTRONIX Field Office or representative in your area. This will assure you the fastest possible service. Please include the instrument Type Number or Part Number and Serial Number with all requests for parts or service.

Specifications and price change privileges reserved.

Copyright © 1974 by Tektronix, Inc., Beaverton, Oregon. Printed in the United States of America. All rights reserved. Contents of this publication may not be reproduced in any form without permission of Tektronix, Inc.

U.S.A. and foreign TEKTRONIX products covered by U.S. and foreign patents and/or patents pending.

TEKTRONIX is a registered trademark of Tektronix, Inc.

TABLE OF CONTENTS

BEFORE OPERATING	Page
INTRODUCTION	1
SAFETY INFORMATION	1
OPERATING VOLTAGE	2
OPTIONS	3
CONTROLS, CONNECTORS AND INDICATORS	
VERTICAL	4
DISPLAY	6
DM43/DM40	6
TRIGGER	8
HORIZONTAL, CALIBRATOR, AND POWER	12
REAR PANEL	14
SIDE PANEL	15
BASIC OSCILLOSCOPE DISPLAYS	
NORMAL SWEEP DISPLAY	16
MAGNIFIED SWEEP DISPLAY	17

BASIC OSCILLOSCOPE DISPLAYS (cont)	Page
DELAYED SWEEP DISPLAY	17
MIXED SWEEP DISPLAY	18
X-Y DISPLAY	18
SINGLE SWEEP DISPLAY	19
DM43/DM40 DISPLAYS AND MEASUREMENTS	
RESISTANCE	20
VOLTS	22
TEMPERATURE (DM43 ONLY)	24
APPLICATIONS AND MEASUREMENTS	
PRELIMINARY	28
OPERATOR'S ADJUSTMENTS AND CHECKS	29
PEAK-TO-PEAK VOLTAGE MEASUREMENTS —AC	31
INSTANTANEOUS VOLTAGE MEASUREMENTS —DC	32
ALGEBRAIC ADDITION	34

TABLE OF CONTENTS (cont)

APPLICATIONS AND MEASUREMENTS (cont)	Page
COMMON-MODE REJECTION	35
AMPLITUDE COMPARISON MEASUREMENTS	36
TIME DURATION MEASUREMENTS	37
FREQUENCY MEASUREMENTS	38
RISETIME MEASUREMENTS	39
TIME-DIFFERENCE BETWEEN TWO PULSES FROM DIFFERENT SOURCES	40
TIME COMPARISON MEASUREMENTS	41
BASIC 465 DELAYED OR MIXED SWEEP TIME MEASUREMENTS	42
Time Difference Between Repetitive Pulses (Basic 465)	42
Time Duration Measurements (Basic 465)	43
Frequency (Basic 465)	44
Time Difference Between Two Pulses From Different Sources (Basic 465)	45
Risetime (Basic 465)	47
Mix (Basic 465)	48

APPLICATIONS AND MEASUREMENTS (cont)	Page
DM43/DM40 DELAYED OR MIXED SWEEP TIME MEASUREMENTS	48
Time Difference Between Repetitive Pulses (DM43/DM40)	48
Time Duration (DM43/DM40)	49
Frequency (DM43/DM40)	50
Time Difference Between Two Pulses From Different Sources (DM43/DM40)	51
Risetime (DM43/DM40)	52
Mix (DM43/DM40)	53
DELAYED OR MIXED SWEEP MAGNIFICATION	53
Magnified Sweep Starts After Delay	53
Triggered Delayed Sweep Magnification	55
PHASE DIFFERENCE MEASUREMENTS	56
HIGH RESOLUTION PHASE DIFFERENCE MEASUREMENTS	57
PULSE JITTER MEASUREMENTS	58

TABLE OF CONTENTS (cont)

OPERATOR'S SPECIFICATIONS	Page
VERTICAL	59
TRIGGERING	60
DIFFERENTIAL TIME MEASUREMENT (BASIC 465)	60
DIFFERENTIAL TIME MEASUREMENT (DM43/DM40)	61
HORIZONTAL	61
CALIBRATOR	61
Z-AXIS INPUT	62
OUTPUTS	62
RESISTANCE DM43/DM40	62
TIME DM43/DM40	62

OPERATOR'S SPECIFICATIONS (cont)	Page
TEMPERATURE—DM43 ONLY, WITH PROBE	63
DC VOLTAGE DM43/DM40	63
AC POWER SOURCE	64
ENVIRONMENTAL	64
ACCESSORIES	
STANDARD ACCESSORIES INCLUDED	65
OPTIONAL ACCESSORIES	65
OPTIONS	
OPTION 4	66
OPTION 7	66



Basic 465 Oscilloscope.



465 Oscilloscope with DM 43 digital multimeter.

BEFORE OPERATING

INTRODUCTION

The Tektronix 465 Oscilloscope is a dual-channel, portable instrument. The dual-channel dc-to-100 MHz vertical system provides calibrated deflection factors from 5 millivolts to 5 volts/division. The sweep trigger circuits are capable of stable triggering over the full bandwidth capabilities of the vertical deflection system. The horizontal deflection system provides calibrated sweep rates from 0.5 second to 0.05 microsecond/division along with delayed sweep features for accurate relative-time measurements. A X10 magnifier extends the calibrated sweep rate to 5 nanoseconds/division. The instrument operates over a wide variation of line voltages and frequencies. Maximum power consumption is about 100 watts.

The Tektronix DM43 Digital Mulitmeter measures 0 ohms to 20 megohms, 0 dc volts to 1200 dc volts (+ or −) or −55° C to +150° C and displays the measurement on a digital display while the oscilloscope operates normally.

The digital multimeter and oscilloscope combine to provide a digital readout of the time between any two points on the oscilloscope display.

The DM40 is identical to the DM43, except that the DM40 lacks temperature measurement capability.

SAFETY INFORMATION

The instrument is designed to operate from a single-phase power source with one of the current-carrying conductors (the Neutral Conductor) at ground (earth) potential. Operation from power sources where both current-carrying conductors are live with respect to ground (such as phase-to-phase on a three-wire system) is not recommended, since only the Line Conductor has over-current (fuse) protection within the instrument.

The instrument has a three-wire power cord with a three-terminal polarized plug for connection to the power source and safety-earth. The ground (earth) terminal of the plug is directly connected to the instrument frame. For electric-shock protection, insert this plug only in a mating outlet with a safety-earth contact.

OPERATING VOLTAGE

CAUTION

This instrument may be damaged if operated with the line voltage selector switch or the regulating range selector set for the wrong applied line voltage or if the wrong line fuse is used.

Line Voltage Selection

This instrument operates from either a 115-volt or a 230-volt nominal line voltage source, 48 to 440 hertz. To convert from one nominal line voltage range to the other, move the Line Voltage Selector switch (located on side panel) to indicate the correct nominal voltage. A 115-to-230 volt adapter may be required for the line-cord plug. Be sure to use the correct line fuse for the line voltage selected (see Table 1).

Regulating Range Selection

The Regulating Range Selector assembly (located on the rear panel) is set for one of the line voltage ranges shown in Table 1. It also contains the line fuse for overload protection.

To change the regulating range:

1. Disconnect the instrument from the power source.

2. Loosen the two captive screws that hold the cover on the selector assembly; then pull to remove the cover.

3. Pull out the range selector switch bar (see Fig. 1). Select a range from Table 1 which is centered about the average line voltage. Slide the bar to the desired position and plug it in. Push the cover on and tighten the screws.

TABLE 1

Regulating Ranges

	Regulating Range	
Range Selector Switch Position	**115-Volts Nominal**	**230-Volts Nominal**
LO (Switch bar in lower holes)	99 to 121 volts	198 to 242 volts
M (Switch bar in middle holes)	104 to 126 volts	207 to 253 volts
HI (Switch bar in upper holes)	108 to 132 volts	216 to 264 volts
Fuse Size	1.5 A 3AG Fast-blow	0.75 A 3AG Fast-blow

Fig. 1. Regulating range selector and line fuse.

OPTIONS

Options are available to alter oscilloscope performance to meet particular applications. A number in either MOD slot (see Item 52, Controls, Connectors, and Indicators) indicates a modified oscilloscope.



Fig. 2. Vertical controls, connectors, and indicators.

CONTROLS, CONNECTORS AND INDICATORS

VERTICAL

1. CH 1 and CH 2 VOLTS/DIV — Selects the vertical deflection factor in a 1-2-5 sequence (VAR control must be in the calibrated detent for the indicated deflection factor).

2. VOLTS/DIV Readout — Consists of two small lamps for each channel, located beneath the skirt of each VOLTS/DIV knob. One or the other lamp will light up to indicate the correct deflection factor when a probe with a scale-switching connector is used. A probe without this connector lights the X1 lamp.

3. VAR — Provides continuously variable uncalibrated deflection factors between the calibrated settings of the VOLTS/DIV switch, and extends the maximum vertical deflection to at least 12.5 volts per division (5 volt position).

4. UNCAL Lamp — Indicates when the VAR VOLTS/DIV control is out of the calibrated detent and the vertical deflection factor is uncalibrated.

5. POSITION — Controls the vertical position of the trace. In the X-Y mode, the CH 2 POSITION control positions on the Y-axis (vertically) and the Horizontal POSITION control positions on the X-axis (horizontally).

6. CH 1 OR X and CH 2 OR Y — Input connectors for application of external signals to the inputs of the vertical amplifier. In the X-Y mode of operation, the signal connected to the CH 1 OR X connector provides horizontal deflection and the signal connected to the CH 2 OR Y connector provides vertical deflection.

7. AC-GND-DC — Selects the method used to couple a signal to the input of the vertical amplifier. In the AC position, signals are capacitively coupled to the vertical amplifier. The dc component of the input signal is blocked. In the GND position, the input of the vertical amplifier is disconnected from the input connector and grounded to allow the input coupling capacitor to precharge. In the DC position, all components of the input signal are passed to the input amplifier.

8. VERT MODE — Selects mode of operation for vertical amplifier system.

CH 1: Channel 1 only is displayed.

ALT: Provides dual-trace display of the signals of both channels. Display is switched between channels at the end of each sweep. Useful at sweep rates faster than about 50 microseconds/division.

ADD: Signals applied to the CH 1 and CH 2 input connectors are algebraically added, and the algebraic sum is displayed on the crt. The INVERT switch in Channel 2 allows the display to be CH 1 plus CH 2 or CH 1 minus CH 2. Useful for common-mode rejection to remove an undesired signal or for dc offset.

CHOP: Provides dual-trace display of the signals of both channels. Display is switched between channels at a repetition rate of approximately 250 kHz. Useful at sweep rates slower than about 50 microseconds/division, or when a dual-trace, single-sweep display is required.

CH 2: Channel 2 only is displayed. It must be selected in X-Y operation.

9. 20 MHz BW/TRIG VIEW — Dual-purpose switch that limits the bandwidth of the vertical amplifier system to approximately 20 MHz when pulled, or when pressed, causes the signal applied to A Trigger Generator to be displayed on the crt.

10. INVERT — Channel 2 display is inverted in the INVERT (button in) position.



Fig. 3. Display and DM43/DM40 controls, connectors, and indicators.

DISPLAY

11. Internal Graticule — Eliminates parallax. Risetime, amplitude and measurement points are indicated at the left-hand graticule edge.

12. BEAM FINDER — Compresses the display to within the graticule area independently of display position or applied signals and provides a visible viewing level.

13. INTENSITY — Controls brightness of the crt display.

14. FOCUS — Adjusts for optimum display definition.

15. SCALE ILLUM — Controls graticule illumination.

16. ASTIG — Used in conjunction with the FOCUS control to obtain a well-defined display. It does not require readjustment in normal use.

17. TRACE ROTATION — Adjusts trace to align with the horizontal graticule lines.

DM43/DM40

18. Input Connectors — Two banana jacks provide + (red) and COM (black) inputs for voltage and resistance only.

19. Probe Connector — Deleted on DM40. Input connector for the temperature probe only.

20. Readout — Negative polarity indication is automatic for negative dc voltage and temperature with no polarity indication for positive measurement. A blinking display indicates overrange. The decimal point is positioned by the FUNCTION and RANGE controls for multimeter operation and the oscilloscope A TIME/DIV switch in the TIME mode.

21. RANGE — Selects from .2 V to 1.2 kV dc in 5 ranges (1200 volts is the maximum safe input in the 1.2 kV dc mode); from 200 ohms to 20 megohms in 6 ranges.

22. FUNCTION — Selects VOLTS, OHMS, TIME or TEMP (°C), except the DM40 lacks the TEMP (°C) function.

23. TIME ms and μs Lamps — Indicates measurement time. There is a millisecond lamp and a microsecond lamp; no lamp lit indicates seconds.

24. TIME ZERO — Resets the TIME readout to zero at any reference point on the oscilloscope 'A' sweep in the TIME mode.



Fig. 4. Trigger controls, connectors, and indicators.

TRIGGER

25. TRIG MODE — Determines the mode of trigger operation for A Sweep.

AUTO: Sweep is initiated by the applied trigger signal. In the absence of an adequate trigger signal, or if the trigger repetition rate is less than about 20 hertz, the sweep free runs and provides a bright reference trace.

NORM: Sweep is initiated by the applied trigger signal. In the absence of an adequate trigger signal, there is no trace. When the trigger rate is too low for AUTO use NORM.

SINGL SWP: When this pushbutton is pushed, the A Sweep operates in the single sweep mode. After a single sweep is displayed, further sweeps cannot be presented until the SINGL SWP button is again pushed. It is useful when the signal to be displayed is not repetitive or varies in amplitude, shape or time causing an unstable conventional display. It can also be used to photograph a nonrepetitive signal.

26. READY Lamp — Indicates A Sweep is "armed" and, upon receipt of an adequate trigger signal, will present a single-sweep display.

27. TRIG Lamp — Indicates that A Sweep is triggered and will produce a stable display. It is useful for setting up the trigger circuits when a trigger signal is available without a display on the crt (for example, when using external triggers).

28. A TRIG HOLDOFF — Provides continuous control of time between sweeps. Allows triggering on aperiodic signals (such as complex digital words). In the fully clockwise position (B ENDS A), A sweep is reset at the end of B sweep to provide the fastest possible sweep repetition rate for delayed-sweep presentations and low-repetition rate signals. Use the A trigger controls for the best possible display before using the A TRIG HOLDOFF control.

29. COUPLING — Determines method used to couple signals to the trigger generator circuit.

AC: Signals are capacitively coupled to the input of the trigger generator. Dc is rejected and signals below about 30 Hz are attenuated.

LF REJ: Signals are capacitively coupled to the input of the trigger circuit. Dc is rejected and signals below about 50 kHz are attenuated. It is useful for providing a stable display of the high-frequency components of a complex waveform.

HF REJ: Signals are capacitively coupled to the input of the trigger circuit. Dc is rejected and signals below about 30 Hz and above about 50 kHz are attenuated. It is useful for providing a stable display of the low-frequency components of a complex waveform.

DC: All components of a trigger signal are coupled to the input of the trigger circuit. It is useful for providing a stable display of low-frequency or low-repetition rate signals, except the combination of the ALT (dual trace) mode with the trigger SOURCE switch in NORM.



Fig. 5. Trigger controls, connectors, and indicators.

30. SLOPE — Selects the slope of the trigger signal that starts the sweep.

+: Sweep can be triggered from the positive-going portion of a trigger signal.

−: Sweep can be triggered from the negative-going portion of a trigger signal.

Correct SLOPE setting is important in obtaining a display when only a portion of a cycle is being displayed.

31. LEVEL — Selects the amplitude point on the trigger signal at which the sweep is triggered. It is usually adjusted for the desired display after trigger SOURCE, COUPLING and SLOPE have been selected.

32. SOURCE — Determines the source of the trigger signal coupled to the input of the trigger circuit.

NORM: Trigger source is displayed signal(s). It does not indicate time relationship betwen CH 1 and CH 2 signals. However, stable triggering of nontime related signals usually can be obtained by setting VERT MODE to ALT, SOURCE to NORM and COUPLING to LF REJ. Carefully adjust LEVEL for a stable display.

CH 1: A sample of the signal available in Channel 1 is used as a trigger signal. CH 2 signal is unstable if it is not time related.

CH 2: A sample of the signal available in Channel 2 is used as a trigger signal. CH 1 signal is unstable if it is not time related.

LINE (A Trigger circuit only): A sample of the power-line frequency is used as a trigger signal. It is useful when the input signal is time-related (multiple or submultiple) to the line frequency or when it is desirable to provide a stable display of a line-frequency component in a complex waveform.

EXT: Signals connected to the External Trigger Input connectors are used for triggering. External signals must be time-related to the displayed signal for a stable display. It is useful when the internal signal is too small or contains undesired signals that could cause unstable triggering. It is useful when operating in CHOP Mode.

EXT ÷ 10 (A Trigger circuit only): External trigger signal is attenuated by a factor of 10.

STARTS AFTER DELAY (B Trigger circuit only): B Sweep runs immediately after the delay time selected by the DELAY-TIME POSITION control.

33. External Trigger Inputs — Input connectors for external trigger signals.



Fig. 6. Horizontal, calibrator, and power.

HORIZONTAL, CALIBRATOR, AND POWER

34. FINE/POSITION — Positions the display horizontally for A Sweep, B Sweep or the X-axis in the X-Y mode.

35. X10 MAG — Increases displayed sweep rate by a factor of 10. It extends fastest sweep rate to 5 nanoseconds/division. The magnified sweep is the center division of the unmagnified display (0.5 division either side of the center graticule line).

36. HORIZ DISPLAY — Determines the mode of operation for the horizontal deflection system.

A: Horizontal deflection provided by A Sweep at a sweep rate determined by the setting of the A TIME/DIV switch. B sweep is inoperative.

MIX: The first part of the horizontal sweep is displayed at a rate set by the A TIME/DIV switch and the latter part of the sweep at a rate set by the B TIME/DIV switch. Relative amounts of the display allocated to each of the two sweep rates are determined by the setting of the DELAY-TIME POSITION control.

A INTEN: The sweep rate is determined by the A TIME/DIV switch. An intensified portion appears on the display during the B Sweep time, which is about 10 times

the B TIME/DIV switch setting. This switch position provides a check of the duration and position of the B Sweep (delayed sweep) with respect to the A Sweep (delaying sweep).

B DLYD: The sweep rate is determined by the B TIME/DIV switch with the delay time determined by the setting of the DELAY TIME (A TIME/DIV) switch and the DELAY-TIME POSITION control.

37a. A AND B TIME/DIV AND DELAY TIME — A TIME/DIV switch (clear plastic skirt) selects the sweep rate of the A Sweep circuit for A sweep only operation and selects the basic delay time (to be used by the DELAY-TIME POSITION control) for delayed sweep operation. B TIME/DIV switch (pull out and rotate to unlock) selects the sweep rate for the B Sweep circuit for delayed sweep operation only. A VAR control must be in the calibrated detent for calibrated A Sweep rates. When both TIME/DIV switches are fully counterclockwise to the X-Y position, the horizontal (X-axis) display is the CH 1 input.

37b. DM43/DM40 A AND B TIME/DIV AND DELAY TIME — The controls operate in the same manner as 37a. above. The A TIME/DIV switch also controls the TIME lamps and decimal point location when the DM43/DM40 is in the TIME mode.

38. VAR — Provides continuously variable sweep rates between the calibrated settings of the A TIME/DIV switch. It extends the slowest A Sweep rate to at least 1.25 seconds/division. The A Sweep rate is calibrated when the control is set fully clockwise to the calibrated detent.

39. UNCAL Lamp — Indicates the A Sweep rate is uncalibrated (VAR control out of the calibrated detent).

40. X10 MAG Lamp — Indicates that the X10 magnifier is on.

41a. DELAY TIME POSITION — Provides variable sweep delay to more than 10 times the delay time indicated by the A TIME/DIV switch.

41b. DM43/DM40 DELAY TIME POSITION — The control operates in the same manner as 41a. above. When the DM43 or DM40 is in the TIME mode, the DELAY-TIME POSITION control, in conjunction with the ZERO reset button, measures time between any two points on the oscilloscope display.

42. CALIBRATOR — A combination current loop/square-wave voltage output that permits the operator to compensate voltage probes and check vertical gain, current probes and oscilloscope operation. It is not intended to verify time-base calibration.

43. POWER — Turns instrument power on and off.

44. LOW LINE Lamp — Indicates that the applied line voltage is below the lower limit of the regulating range selected by the Regulating Range Selector.



Fig. 7. Rear panel controls, connectors, and indicators.

REAR PANEL

45. A +GATE — Provides a positive-going rectangular pulse coincident with the A Sweep time, which can be used to trigger the signal source (with TRIG MODE switch set to AUTO).

46. B +GATE — Provides a positive-going rectangular pulse coincident with the B Sweep time, which can be used to trigger the signal source after a selected delay time, providing that A Sweep is triggered internally.

47. CH 1 VERT SIGNAL OUT — Provides a sample of the signal applied to the CH 1 input connector.

48. EXT Z-AXIS — Permits intensity modulation of the crt display. Does not affect display waveshape. Signals with fast rise and fall provide the most abrupt intensity change. Signals must be time-related to the display for a stable display. Useful for adding time markers in uncalibrated modes of operation.

49. Line Fuse Holder — Contains the line fuse and the regulating range selector. See Fig. 1 for change information.

50. Regulating Range Selector — Shown in Medium regulating range. See Fig. 1 for change information.

51. Line Cord — Makes the connection between the oscilloscope and the power source. The cord may be conveniently stored by wrapping it around the feet on the rear panel.

52. MOD SLOTS — A number in either slot indicates the instrument contains an option or other modifications.

SIDE PANEL

53. LINE VOLTAGE SELECTOR SWITCH (located on the right-hand side) — Selects either 115 V or 230 V nominal line voltage.

BASIC OSCILLOSCOPE DISPLAYS

These instructions permit the operator to obtain the most commonly used basic displays.

NORMAL SWEEP DISPLAY

1. Set the controls as follows:

Vertical

VERT MODE	CH 1
VOLTS/DIV	Position determined by amplitude of signal to be applied.
VOLTS/DIV VAR	Calibrated detent.
AC—GND—DC	AC
Vertical POSITION	Midrange.
20 MHz BW	Not limited (yellow band not visible).
INVERT	Button out.

Horizontal

TIME/DIV Switches	Locked together at 1 ms.
A TIME/DIV VAR	Calibrated detent.
HORIZ DISPLAY	A
X10 MAG	Off (button out)
POSITION	Midrange

Display

INTENSITY	Fully counterclockwise.

Trigger

(Both A and B if applicable)

SLOPE	+.
LEVEL	0.
SOURCE	NORM.
COUPLING	AC.
TRIG MODE (A only)	AUTO.
A TRIG HOLDOFF	NORM.

2. Pull the POWER switch (on). Connect the signal to the CH 1 input connector.

3. Adjust the INTENSITY control for the desired display brightness. If the display is not visible with the INTENSITY control at midrange, press the BEAM FINDER pushbutton and adjust the CH 1 VOLTS/DIV switch to reduce the vertical display size. Center the compressed display with the vertical and horizontal POSITION controls; release the BEAM FINDER pushbutton.

4. Set the CH 1 VOLTS/DIV switch and vertical POSITION control to locate the display within the display area.

5. Adjust the A Trigger LEVEL control for a stable display.

6. Set the A TIME/DIV switch and the horizontal POSITION control to locate the display within the display area. Then adjust the FOCUS control as needed.

MAGNIFIED SWEEP DISPLAY

1. Obtain a Normal Sweep Display.

2. Adjust the horizontal POSITION control to move the area to be magnified to within the center graticule division (0.5 division on each side of the center vertical graticule line). It may be necessary to change the TIME/DIV switch setting.

3. Push the X10 MAG switch (on) and adjust the horizontal POSITION control for precise positioning of the magnified display. Divide the TIME/DIV setting by 10 to determine the magnified sweep rate.

DELAYED SWEEP DISPLAY

1. Obtain a Normal Sweep Display.

2. Set the HORIZ DISPLAY switch to A INTEN and the B Trigger SOURCE switch to STARTS AFTER DELAY.

3. Pull out the B TIME/DIV switch knob and turn clockwise so the intensified zone on the display is the desired length. Adjust the INTENSITY control for the desired display brightness.

4. Adjust the DELAY-TIME POSITION control to move the intensified zone to the portion of the display to be delayed.

5. Set the HORIZ DISPLAY switch to B DLY'D. The intensified zone on the display noted in steps 3 and 4 is now displayed in delayed form. The delayed sweep rate is indicated by the dot on the B TIME/DIV switch knob.

6. For a delayed sweep display with less jitter, set the B Trigger SOURCE switch to the same position as the A Trigger SOURCE switch and adjust the B Trigger LEVEL control for a stable display. If the A Trigger SOURCE switch is in the LINE position, a sample of the line voltage will have to be supplied to the B Trigger circuit externally via the B Trigger external trigger connector.

MIXED SWEEP DISPLAY

1. Obtain a Normal Sweep Display.

2. Pull out the B TIME/DIV switch knob and turn clockwise to the desired sweep rate. Adjust the INTENSITY control for the desired display brightness.

3. Set the HORIZ DISPLAY switch to MIX. The display now contains more than one time factor on the horizontal axis. The first portion of the display is at the A Time Base sweep rate and the latter part is at the B Time Base sweep rate. The start of the B Time Base portion of the display can be changed by adjusting the DELAY-TIME POSITION control.

X-Y DISPLAY

1. Preset the instrument controls as given in step 1 of Normal Sweep Display, then turn the instrument power on. Allow several minutes for instrument warm-up.

2. Set the TIME/DIV switches to X-Y and the VERT MODE switch to CH 2. Apply the vertical signal to the CH 2 OR Y input connector and the horizontal signal to the CH 1 OR X input connector.

3. Advance the INTENSITY control until the display is visible. If the display is not visible with the INTENSITY

control at midrange, press the BEAM FINDER pushbutton and adjust the CH 1 and CH 2 VOLTS/DIV switches until the display is reduced in size, both vertically and horizontally. Center the compressed display with the POSITION controls (CH 2 POSITION vertical, CH 1 POSITION/FINE horizontally); release the BEAM FINDER pushbutton. Adjust the FOCUS control for a well-defined display.

SINGLE SWEEP DISPLAY

1. Obtain a Normal Sweep Display. For random signals, set the trigger circuit to trigger on a signal that is approximately the same amplitude and frequency as the random signal.

2. Push the SINGL SWP button on the A TRIG MODE switch. The next trigger pulse starts the sweep and displays a single trace. If no triggers are present, the READY lamp lights, indicating the A Sweep Generator circuit is set and waiting to be triggered.

3. After the sweep is complete, the circuit is "locked out" and the READY lamp is out. Press the SINGL SWP button to prepare the circuit for another single-sweep display.

DM43/DM40 DISPLAYS AND MEASUREMENTS

Except for the time function, the DM43 (or DM40) is independently usable anytime the oscilloscope is turned on. The time function is discussed in the Oscilloscope Applications and Measurements section of this manual under Delayed or Mixed Sweep Measurements.

RESISTANCE

CAUTION

The meter may be damaged by attempting to measure voltage if the meter is in the resistance mode of operation (OHMS FUNCTION button pushed in) and the applied voltage is in excess of 120 V rms.

1. Push in the OHMS FUNCTION button and the 20 MΩ RANGE button. See Fig. 8.

2. Connect the + and COM leads to the unknown resistance.

3. Reduce the range, using the following table, until a proper readout is obtained.

TABLE 2

Resistance Ranges

RANGE	READOUT	MEASUREMENT
20 MΩ	20.00—02.00	20 MΩ—2MΩ
2 MΩ	2.000—0.200	2 MΩ—200 kΩ
200 kΩ	200.0—20.00	200 kΩ—20 kΩ
20 kΩ	20.00—02.00	20 kΩ—2 kΩ
2 kΩ	2.000—0.200	2 kΩ—200 Ω
200 Ω	200.0—000.0	200 Ω—0 Ω

A blinking readout on any range, when connected to any unknown resistance, indicates an overrange condition and a need to select a higher range.

A blinking readout, when meter leads are disconnected is normal.



Fig. 8. Resistance.

VOLTS

CAUTION

The maximum safe input voltage is ±500 V (dc + peak ac) for the 200 mV and 2 V ranges, and ±1200 V (dc + peak ac) for the 20 V, 200 V and 1.2 kV ranges; between the + and COM inputs or between the + input and chassis.

The maximum COM floating voltage is ±500 V (dc + peak ac) to chassis.

The meter may be damaged by attempting to measure voltage if the meter is in the resistance mode of operation and the applied voltage is in excess of 120 V rms.

If the reading exceeds 1200 V or the readout blinks (indicating overrange) disconnect the + lead at once to prevent possible meter damage.

1. Push in the VOLTS FUNCTION button and the 1.2 kV RANGE button. See Fig. 9.

2. Connect the COMMON lead to the reference point (usually a ground or test point) and the HIGH lead to the unknown voltage to be measured and observe the reading.

3. Reduce the range, using the following table, until a proper readout is obtained.

NOTE

When no voltage is applied in the 20 V to 1.2 kV ranges, the readout is 0000 and individual readout elements may blink. Also, noise picked up by the meter leads may increase the readout in the .2 V and 2 V ranges.

A blinking readout on any range indicates an overrange condition and a need to select a higher range.

TABLE 3

Volts Ranges

RANGE	READOUT	MEASUREMENT
1.2 kV	1.200—0.200	1.2 kV—200 V
200 V	200.0—020.0	200 V—20 V
20 V	20.000—02.00	20 V —2 V
2 V	2.000—0.200	2V—.2V
200 mV	0.200—0.000	.2V —0 V


+ (RED)
COM (BLACK)
VOLTAGE INPUT
REFERENCE
FUNCTION SELECTION
RANGE SELECTION
465/DM-O-3

Fig. 9. Volts.

TEMPERATURE (For DM43 only)

CAUTION

The maximum safe voltage on the measurement surface is ±400 V (dc + peak ac) above chassis ground.

The sensor tip is fragile and may break if dropped or subjected to excessive stress. Force exerted on the sensor tip should not exceed 20 pounds.

If the reading exceeds −55° C or +150° C, remove the probe at once to prevent probe damage.

1. Push in the TEMP (° C) FUNCTION button. See Fig. 10.

2. Apply the temperature probe to the device to be measured so the flat surface of the probe tip mates against the device to be measured and observe the reading. See Table 4 to convert the reading from centigrade to fahrenheit.

NOTE

Temperature accuracy is affected by several factors such as the angle of contact between the probe and the device to be measured, a thermal gradient, heat removed from the device by the probe, radio-frequency signals present, etc.

These items and other probe information are in the probe manual, which should be reviewed.

Accuracy Check

The DM43 is calibrated to its original probe, giving accurate readings within 2° C for examples that follow. The DM43 should be recalibrated to any replacement probe.

In the following checks, use an accurate thermometer to verify water temperature. Anything in solution affects the melting temperature and the boiling point is affected by changes in altitude and barometric pressure.

Low Temperature

Allow a container (preferably insulated) of crushed ice to melt until there are only a few pieces of ice remaining.



Fig. 10. Temperature.

TABLE 4
CENTIGRADE TO FAHRENHEIT CONVERSION



Put the probe tip into the water, avoiding the side or bottom of the container. Wait for the readout to stabilize, indicating the probe has reached the water temperature.

The readout should be −2 to 2. There should be ice remaining after the test to verify that inserting the probe did not raise the water temperature.

High Temperature

Bring water to a slow boil (to prevent splattering). Put the probe tip into the water, avoiding the side or bottom of the container. Wait for the readout to stabilize, indicating the probe has reached the water temperature.

The readout should be between 98 and 102 for clean water at sea level.

APPLICATIONS AND MEASUREMENTS

Refer to Basic Oscilloscope Displays as needed to obtain correct displays.

PRELIMINARY

Signal Ground

The most reliable signal measurements are made when the oscilloscope and the unit under test are connected togther by a common reference (ground) lead in addition to the signal lead or probe. The ground strap on the signal probe provides the best ground. Also, a ground lead can be connected to the ground banana jack on the oscilloscope chassis to establish a common ground with the signal source.

Input Coupling Capacitor Precharging

In the GND position, the input signal is connected to ground through a one-megohm resistor to form a precharging network. This network allows the Input Coupling capacitor to charge to the average dc voltage level of the signal applied to the probe. Thus, any large voltage transients accidentally generated will not be applied to the amplifier input. The precharge network also provides a measure of protection to the external circuitry by reducing the current levels that can be drawn from the external circuitry during capacitor charging.

The following procedure should be used whenver the probe tip is connected to a signal source having a different dc level than that previously applied.

1. Before connecting the probe tip to a signal source, set the Input Coupling switch to GND.

2. Touch the probe tip to oscilloscope chassis ground. Wait several seconds for the Input Coupling capacitor to discharge.

3. Connect the probe tip to the signal source.

4. Wait several seconds for the Input Coupling capacitor to charge.

5. Set the Input Coupling Switch to AC. The display will remain on screen so the ac component of the signal can be measured in the normal manner.

OPERATOR'S ADJUSTMENT AND CHECKS

To verify measurement accuracy, perform the following checks and adjustments before making a measurement. See the Calibration section of the Service Manual for calibration information.

Trace Rotation Adjustment

(Normally not required.) Obtain a Normal Sweep Display using only steps 1 through 3. Set the CH 1 Input Coupling switch to GND to display a free-running trace with no vertical deflection. Adjust the TRACE ROTATION (screwdriver adjustment located below the crt graticule) to align the trace with the center horizontal graticule line.

Probe Compensation

(One of the greatest sources of operator error.) Most attenuator probes are equipped with adjustments to ensure optimum measurement accuracy. Probe compensation is accomplished as follows:

Obtain a Normal Sweep Display presentation of the calibrator square-wave voltage. Set the appropriate VOLTS/DIV switch to .1 V position and the Input Coupling to DC. Check the waveform presentation for overshoot or rolloff, and if necessary, readjust compensation for flat tops on the waveforms. See Fig. 11. Refer to probe manual for method of compensating the probe being used.



Fig. 11. Probe compensation.

Vertical Gain Check

Obtain a Normal Sweep Display presentation of the calibrator square-wave voltage. Set the appropriate VOLTS/DIV switch to the 50 mV position and the Input Coupling switch to DC. Make sure the VAR VOLTS/DIV control is in the calibrated detent. Check that the vertical deflection is 5.8 to 6.2 divisions.

Basic 465 Timing Check (60 Hertz Line Only)

Obtain a Normal Display. Set the A TIME/DIV switch to the 5 ms position. Set the A Trigger SOURCE switch to LINE. Push the TRIG VIEW switch and hold it in. This displays a sample of the line voltage. Use the A Trigger LEVEL control to vertically position the top of the display to within the display area. Use the horizontal position control to position the left peak to the left graticule edge (see Fig. 12). Verify the horizontal distance between the first and the fourth peaks is 9.8 to 10.2 divisions. If the fourth peak is not visible, verify the horizontal distance between the first and the third peaks is 6.53 to 6.79 divisions.

DM43/DM40 Timing Check (60 Hertz Line Only)

Do the Basic 465 Timing Check given in the previous step.



Fig. 12. Basic 465 timing check.

Leave the controls as for the Basic 465 Timing Check except as follows. Set the B TIME/DIV switch to 5 μs. Set the HORIZ DISPLAY switch to A INTEN. Set the B Trigger SOURCE switch to STARTS AFTER DELAY. Adjust the INTENSITY control for a visible intensified zone.

Push the TRIG VIEW switch and hold it in to obtain the same display as for Basic 465 Timing Check. Use the DELAY TIME POSITION control to move the intensified zone so it starts at the first positive going portion that crosses the center horizontal graticule line (see Fig. 13, Point A). Push the ZERO button on DM43/DM40 to set the readout to zero. Use the DELAY TIME POSITION control to move the intensified zone to the second positive going portion of the display crossing the center horizontal graticule line (see Fig. 13, Point B). Verify that the readout is 16.44 to 16.89 and the ms lamp is lit.



Fig. 13. DM43/DM40 timing check.

External Horizontal Gain Check

(If X-Y operation is to be used.) Use steps 1 through 3 of the Basic Oscilloscope Displays procedure for obtaining a Normal Sweep Display of the calibrator square-wave voltage waveform; then, set the TIME/DIV switch to X-Y. With the calibrator signal connected to the CH 1 OR X input connector and the CH 1 VOLTS/DIV switch set to 50 mV, the crt display should be two dots separated horizonally by 5.75 to 6.25 divisions.

PEAK-TO-PEAK VOLTAGE MEASUREMENTS—AC

Obtain a Normal Display. Make sure the VAR VOLTS/DIV control is in the calibrated detent. Vertically position the display so the lower portion coincides with a horizontal graticule line (see Fig. 14 Point A). Horizontally position the display so one of the upper peaks coincides with the center vertical graticule line (see Fig. 14, Point B). Measure the vertical deflection from peak-to-peak (Point A to Point B).



Fig. 14. Peak-to-peak voltage of a waveform.

NOTE

If the amplitude measurement is critical or if the trace is thick as a result of hum and/or noise on the signal, a more accurate measurement can be obtained by measuring from the top of a peak to the top of a valley. This will subtract the trace thickness from the measurement.

Multiply the vertical deflection just measured by the VOLTS/DIV switch setting. Also include the attenuation factor of the probe if the probe does not have a scale-factor switching connector.

EXAMPLE: The peak-to-peak vertical deflection measured is 4.6 divisions (see Fig. 14) with a VOLTS/DIV switch setting of .5.

Using the formula:

$$\text{Volts Peak-to-peak} = \text{vertical deflection (divisions)} \times \text{VOLTS/DIV setting}$$

Substituting the given values:

$$\text{Volts Peak-to-Peak} = 4.6 \times 0.5\ \text{V} = 2.3\ \text{volts.}$$

INSTANTANEOUS VOLTAGE MEASUREMENTS—DC

Obtain a Normal Display. Make sure the VAR VOLTS/DIV control is in the calibrated detent.

To determine the polarity of the voltage to be measured, set the input coupling switch to GND and vertically position the baseline to the center of the crt. Set the input coupling switch to DC. If the waveform moves to above the center of the crt, the voltage is positive. If the waveform moves to below the center of the crt, the voltage is negative.

Set the input coupling switch to GND and position the baseline to a convenient reference line. For example, if the voltage to be measured is positive, then position the baseline to the bottom graticule line.

Switch the Input Coupling Switch to DC. Measure the divisions of vertical deflection between the reference line and the desired point on the waveform (see Fig. 15). Multiply the vertical deflection by the VOLTS/DIV switch setting. Include the attenuation factor of the probe if the probe does not have a scale-factor switching connector.

EXAMPLE: The vertical distance measured is 4.6 divisions (see Fig. 15), the waveform is above the reference line, and the VOLTS/DIV switch is set to 2.



Fig. 15. Instantaneous voltage measurement.

Using the formula:

Instantaneous Voltage	=	vertical distance (divisions)	X	polarity	X	VOLTS/DIV setting

Substituting: = 4.6 X +1 X 2 V = 9.2 volts.

If a negative voltage is to be measured, position the trace to the top graticule line and measure from C to B (see Fig. 15).

The ground reference line can be checked at any time by switching to the GND position.

This method can also be used to measure one voltage with respect to another. Position one of the voltage levels to a convenient reference line. Measure the divisions of vertical deflection between the reference line and the other voltage level. Substitute this value in the formula just given.

ALGEBRAIC ADDITION

In the ADD position of the VERT MODE switch, the waveform displayed is the algebraic sum of the signals applied to the CH 1 and CH 2 inputs (CH 1 + CH 2). If the CH 2 INVERT switch is pushed, the waveform displayed is the difference of the signals applied to the CH 1 and CH 2 inputs (CH 1 − CH 2). The total deflection factor in the ADD mode is equal to the deflection factor indicated by either VOLTS/DIV switch (when both VOLTS/DIV switches are set to the same position).

The following general precautions should be observed when using the ADD mode.

1. Do not exceed the input voltage rating of the oscilloscope.

2. Do not apply signals that exceed an equivalent of about eight times the VOLTS/DIV switch settings. For example, with a VOLTS/DIV switch setting of 0.5, the voltage applied to that channel should not exceed about four volts. Larger voltages may distort the display.

A common use of the ADD mode is to provide a dc offset for a signal riding on a dc level.

EXAMPLE: The Channel 1 signal is on a 3 division, positive dc level (using the center line as zero volts). See Fig. 16A. Multiply 3 divisions by the VOLTS/DIV switch setting to determine the dc-level value. Apply a negative dc level (or a positive level using the CH 2 INVERT switch), of the value determined, to Channel 2 input. See Fig. 16B. The ADD mode puts the resultant display within the operating range of the POSITION controls.



Fig. 16. Algebraic addition.

COMMON-MODE REJECTION

The ADD mode can be used to display signals that contain undesirable components. These undesirable components can be eliminated through common mode rejection. The precautions given under algebraic addition should be observed.

EXAMPLE: The signal applied to the CH 1 input contains unwanted line frequency components (see Fig. 17A.). To remove the undesired components use the following procedure.

1. Connect a line frequency signal to the CH 2 input.

2. Set the VERT MODE switch to ALT and the CH 2 INVERT switch to on (button in). Adjust the CH 2



Fig. 17. Common-mode rejection.

VOLTS/DIV and VAR VOLTS/DIV controls so the CH 2 display is about the same amplitude as the undesired portion of the CH 1 display (see Fig. 17A).

3. Set the VERT MODE switch to ADD. Slightly readjust the CH 2 VAR VOLTS/DIV control for maximum cancellation of the undesired signal component (see Fig. 17B).

AMPLITUDE COMPARISON MEASUREMENTS

If comparisons of an unknown signal with a reference signal are repetitious (e.g., on an assembly line test) it is possible to obtain more accurate easily read measurements if the VAR VOLTS/DIV control is adjusted to set the reference signal to an exact number of divisions. The unknown signal can then be quickly and easily compared with or adjusted to an exact number of divisions.

Other unknown signals may be measured without disturbing the setting of the VAR VOLTS/DIV control by establishing a vertical conversion factor and an arbitrary deflection factor. The amplitude of the reference signal must be known before a vertical conversion factor can be established.

Determine the vertical conversion factor using this formula:

$$\text{Vertical Conversion Factor} = \frac{\text{reference signal amplitude (volts)}}{\text{vertical deflection (divisions)} \times \text{VOLTS/DIV switch setting}}$$

Determine the arbitrary deflection factor using the formula:

$$\text{Arbitrary Deflection Factor} = \text{Vertical Conversion Factor} \times \text{VOLTS/DIV switch setting}$$

To measure the amplitude of an unknown signal, set the VOLTS/DIV switch to a setting that provides sufficient vertical deflection to make an accurate measurement. Do not readjust the VAR VOLTS/DIV control. Measure the vertical deflection in divisions and calculate the amplitude of the unknown signal using the following formula:

$$\text{Signal Amplitude} = \text{Arbitrary deflection factor} \times \text{Vertical deflection (divisions)}$$

or

$$\text{Signal Amplitude} = \text{Vertical Conversion Factor} \times \text{VOLTS/DIV Switch Setting} \times \text{Vertical deflection (divisions)}$$

EXAMPLE: The reference signal amplitude measured is 30 volts with a VOLTS/DIV switch setting of 5 and the VAR VOLTS/DIV control adjusted to provide a vertical deflection of 4 divisions.

Substituting these values in the vertical conversion factor formula:

$$\text{Vertical Conversion Factor} = \frac{30}{4 \times 5\ \text{V}} = 1.5$$

Then with a VOLTS/DIV switch setting of 1, the peak-to-peak amplitude of an unknown signal, 5 divisions high can be determined by using the signal amplitude formula:

Signal Amplitude: = 1 V X 1.5 X 5 = 7.5 volts.

TIME-DURATION MEASUREMENTS

Obtain a Normal Sweep display. Be sure the VAR TIME/DIV control is set to the calibrated detent. Set the TIME/DIV switch for a single event and position the display to place the time measurement points to the center horizontal graticule line (see Fig. 18).



Fig. 18. Time duration.

Measure the horizontal distance between the time measurement points. Multiply the distance measured by the setting of the TIME/DIV switch. If sweep magnification is used, divide this answer by 10.

EXAMPLE: The distance between the time measurement points is 8.3 divisions (see Fig. 18) and the TIME/DIV switch is set to 2 ms with the magnifier off.

Using the formula:

$$\text{Time Duration} = \frac{\text{horizontal distance (divisions)} \times \text{Time/Div setting}}{\text{magnification}}$$

Substitute the given values:

$$\text{Time Duration} = 8.3 \times 2 \text{ ms} = 16.6 \text{ milliseconds}$$

FREQUENCY MEASUREMENTS

Time Duration measurements can be used to determine the frequency of a recurrent signal as follows:

1. Measure one cycle of the waveform as described in Time-Duration Measurements.

2. Take the reciprocal of the time duration to determine the frequency.

EXAMPLE: The frequency of the signal shown in Fig. 18, which has a time duration of 16.6 milliseconds is:

$$\text{Frequency} = \frac{1}{\text{time duration}} = \frac{1}{16.6 \text{ ms}} = 60 \text{ hertz}$$

RISETIME MEASUREMENTS

Risetime measurements use the same methods as time-duration measurements, except the measurements are made between the 10% and 90% points of the waveform. Falltime is measured between the 90% and 10% points on the trailing edge of the waveform.

Obtain a Normal Sweep Display. Use a sweep speed setting that displays several cycles or events (if possible) and be sure the VAR TIME/DIV control is in the calibrated detent. Set the VOLTS/DIV switch and VAR control (or signal amplitude) for exactly a five/division display. Set vertical positioning so the display bottom touches the 0% graticule line and the display top touches the 100% graticule line.

Set the TIME/DIV switch for a single-event display and horizontally position the display so the 10% point of the waveform intersects the second vertical graticule line (see Fig. 19).

Measure the horizontal distance between the 10% and 90% points and multiply the distance measured by the setting of the TIME/DIV switch.



Fig. 19. Risetime.

EXAMPLE: The horizontal distance between the 10% and 90% points is 5 divisions (see Fig. 19) and the TIME/DIV switch is set to 1 μs.

Using the time duration formula to find risetime:

Time Duration (risetime)	=	horizontal distance (divisions)	X	TIME/DIV setting

Substituting the given values:

Risetime = 5 X 1 μs = 5 microseconds

TIME-DIFFERENCE BETWEEN TWO PULSES FROM DIFFERENT SOURCES

Obtain a Normal Sweep Display. Make sure the VAR TIME/DIV control is in the calibrated detent. Set the A Trigger SOURCE switch to CH 1. Connect the reference signal to CH 1 and the comparison signal to CH 2. Connect the signals to the input connectors using probes or cables with equal time delay.

Set the VERT MODE switch to either CHOP or ALT. In general, CHOP is more suitable for low-frequency signals and the ALT position is more suitable for high-frequency signals. Center each of the displays vertically (see Fig. 20).

Measure the horizontal difference between the two signals. Multiply the measured difference by the setting of the TIME/DIV switch; if sweep magnification is used, divide this answer by 10.

EXAMPLE: The TIME/DIV switch is set to 50 μs, the MAG switch to X10 and the horizontal difference between waveforms is 4.5 divisions (see Fig. 20).



Fig. 20. Time difference between two pulses from different sources.

Using the formula:

$$\text{Time Difference} = \frac{\text{Time/Div setting} \times \text{horizontal difference (divisions)}}{\text{magnification}}$$

Substituting the given values:

$$\text{Time Difference} = \frac{50\ \mu\text{s} \times 4.5}{10} = 22.5 \text{ microseconds}$$

TIME COMPARISON MEASUREMENTS

If comparisons of an unknown signal with a reference signal are repetitious (e.g., on an assembly line test) it is possible to obtain more accurate, easily read measurements if the VAR TIME/DIV control is adjusted to set the reference signal to an exact number of divisions. The unknown signal can then be quickly and easily compared with, or adjusted to, an exact number of divisions.

Other unknown signals may be measured without disturbing the setting of the VAR TIME/DIV control by establishing a horizontal conversion factor and an arbitrary deflection factor. The time duration of the reference signal must be known before a horizontal conversion factor can be established.

Determine the horizontal conversion factor using this formula:

$$\text{Horizontal Conversion Factor} = \frac{\text{reference signal time duration (seconds)}}{\text{Horizontal deflection (divisions)} \times \text{VOLTS/DIV switch setting}}$$

Determine the arbitrary deflection factor using this formula:

$$\text{Arbitrary Deflection Factor} = \text{horizontal conversion factor} \times \text{TIME/DIV switch setting}$$

To measure the time duration of an unknown signal, set the TIME/DIV switch to a setting that provides sufficient horizontal deflection to make an accurate measurement. Do not readjust the VAR TIME/DIV control. Measure the horizontal deflection in divisions and calculate the time duration using the formula:

$$\text{Time Duration} = \text{arbitrary deflection factor} \times \text{horizontal deflection (divisions)}$$

or

$$\text{Time Duration} = \text{horizontal conversion factor} \times \text{TIME/DIV switch setting} \times \text{horizontal deflection (divisions)}$$

EXAMPLE: The reference signal frequency measured is 455 hertz (time duration: 2.19 milliseconds) with a TIME/DIV switch setting of .2 ms, and the VAR TIME/DIV control adjusted to provide a horizontal deflection of eight

divisions. Substituting these values in the horizontal conversion factor formula:

$$\text{Horizontal Conversion Factor} = \frac{2.19 \text{ milliseconds}}{0.2 \text{ X } 8} = 1.37$$

Then with a TIME/DIV switch setting of 50 μs, the time duration of an unknown signal that completes one cycle in seven horizontal divisions, can be determined by using the time duration formula:

Time Duration = 50 μs X 1.37 X 7 = 480 μs

This answer can be converted to frequency by taking the reciprocal of the time duration (see applications on Determining Frequency).

BASIC 465 DELAYED OR MIXED SWEEP TIME MEASUREMENTS

The delayed sweep modes can be used to make more accurate time measurements.

Time Difference Between Repetitive Pulses (Basic 465)

Obtain a Delayed Sweep Display. If possible, set the B TIME/DIV switch 20 times faster than the A TIME/DIV switch setting. This produces an intensified zone about 0.5 division in length.

With the HORIZ DISPLAY switch set to A INTEN, use the DELAY TIME POSITION dial to move the intensified zone to the first pulse (see Fig. 21A).

Set the HORIZ DISPLAY switch to B DLY'D. Adjust the DELAY TIME POSITION dial to move the pulse (or rising portion) to some vertical reference line (see Fig. 21B). Note the setting of the DELAY TIME POSITION dial. Turn the DELAY TIME POSITION dial clockwise to move the second pulse to the same vertical reference line (if several pulses are displayed, return to A INTEN to locate the correct pulse). Do not change the settings of the horizontal POSITION and FINE controls. Note the setting of the DELAY TIME POSITION dial.

Determine the time difference using the following formula:

$$\text{Time Difference} = \left[\text{second dial setting} - \text{first dial setting}\right] \text{ X delay time A TIME/DIV switch setting}$$



Fig. 21. Time difference between repetitive pulses.

Example. Assume the first dial setting is 1.31 and the second dial setting is 8.81 with the A TIME/DIV switch set to 0.2 μs (see Fig. 21).

Using the formula:

$$\text{Time Difference} = \left[\begin{array}{c}\text{second dial}\\ \text{setting}\end{array} - \begin{array}{c}\text{first dial}\\ \text{setting}\end{array}\right] \times \begin{array}{c}\text{delay time}\\ \text{(A TIME/DIV}\\ \text{setting)}\end{array}$$

Substituting the given values:

$$\text{Time Difference} = (8.81 - 1.31) \times 0.2\ \mu s.$$

The time difference is 1.5 microseconds.

Time Duration Measurements (Basic 465)

Obtain a Delayed Sweep Display. Set the A TIME/DIV switch to display a single event. Be sure the VAR TIME/DIV control is in the calibrated detent. For the most accurate measurement, set the B TIME/DIV switch to the fastest sweep speed that gives a usable (visible) intensified zone. Vertically position the display to place the time measurement points to the center horizontal graticule line (see Fig. 22).



Fig. 22. Time duration between points on a waveform.

Use the DELAY TIME POSITION dial to move the start (left-hand edge) of the intensified zone to just touch the intersection of the signal and the center horizontal graticule line (see Fig. 22 Point A). Note the DELAY TIME POSITION dial setting.

Use the DELAY TIME POSITION dial to move the start of the intensified zone to the second time measurement point (see Fig. 22 Point B). Note the DELAY TIME POSITION dial setting.

Substitute the DELAY TIME POSITION dial settings into the time difference formula to obtain the time duration.

EXAMPLE: The A TIME/DIV switch is set to 2 ms and the B TIME/DIV switch is set to .1 ms.

The DELAY TIME POSITION dial setting at Point A (Fig. 22) is 1.20. The DELAY TIME POSITION dial at Point B (Fig. 22) is 9.53. To find the time duration use the formula:

$$\text{Time Difference} = \text{Time Duration} =$$

$$\left[\begin{array}{c}\text{second dial setting}\end{array} - \begin{array}{c}\text{first dial setting}\end{array}\right] \begin{array}{c}\text{delay time} \\ \text{X (A TIME/DIV)} \\ \text{(setting)}\end{array}$$

Substituting the given values:

$$\text{Time Duration} = [9.53 - 1.50] \times 2 \text{ ms} = 16.66 \text{ ms}$$

Frequency (Basic 465)

The frequency of a recurrent signal can be calculated by taking the reciprocal of the time duration of one event.

EXAMPLE: The time duration of one event (Point A to Point B, Fig. 22) is 16.66 milliseconds.

Using the formula:

$$\text{Frequency} = \frac{1}{\text{time duration}}$$

Substituting the given values:

$$\text{Frequency} = \frac{1}{16.66\ \text{ms}} = 60\ \text{hertz}$$

Time Difference Between Two Pulses From Different Sources (Basic 465)

Make sure the VAR TIME/DIV control is in the calibrated detent. Set the A Trigger SOURCE switch to CH 1. Connect the reference signal to CH 1 and the comparison signal to CH 2. Connect the signals to the input connectors using probes or cables with equal time delay.

Set the VERT MODE switch to either CHOP or ALT. In general, CHOP is more suitable for low-frequency signals and the ALT position is more suitable for high-frequency signals. Center each of the displays vertically (see Fig. 23A).

Set the HORIZ DISPLAY switch to A INT and the B Trigger SOURCE to STARTS AFTER DELAY. Set the B TIME/DIV switch 20 times faster than the A TIME/DIV switch (when possible).

Use the DELAY TIME POSITION dial to move the intensified zone to the CH 1 pulse. Set the HORIZ DISPLAY switch to B DLY'D. Readjust the DELAY TIME POSITION dial to move the CH 1 pulse or rising portion to some vertical reference line (see Fig. 23B). Note the DELAY TIME POSITION dial setting.

Use the DELAY TIME POSITION dial to move the CH 2 pulse or rising portion to the same reference line. Again note the DELAY TIME POSITION dial setting.

Substitute the DELAY TIME POSITION dial settings in the time difference formula to find the time difference.

EXAMPLE: The A TIME/DIV switch is set to 50 μs and the B TIME/DIV switch is set to 2 μs. Use the DELAY TIME POSITION dial to move the CH 1 pulse to the reference line. The DELAY TIME POSITION dial setting is 2.60.



Fig. 23. Time difference between two pulses from different sources.

Use the DELAY TIME POSITION dial to move the CH 2 pulse to the reference line. The DELAY TIME POSITION dial setting is 7.10.

To find the time difference use the formula:

$$\text{Time difference} = \left[\text{second dial setting} - \text{first dial setting}\right] \text{ delay time X A TIME/DIV setting}$$

Substituting the given values:

$$\text{Time Difference} = [7.10 - 2.60] \times 50\ \mu s = 225\ \mu s$$

Risetime (Basic 465)

Risetime measurements use the same methods as time-duration measurements, except the measurements are made between the 10% and 90% points of the waveform. Falltime is measured between the 90% and 10% points on the trailing edge of the waveform.

Use a sweep speed setting that displays several cycles or events (if possible) and be sure the VAR TIME/DIV control is in the calibrated detent. Set the VOLTS/DIV switch and the VAR control (or signal amplitude) for exactly a five-division display. Set vertical positioning so the display bottom touches the 0% graticule line and the display top touches the 100% graticule line.

Set the A TIME/DIV switch for a single-event display and horizontally position the display so the 10% point of the waveform intersects the second vertical graticule line (see Fig. 24). Set the B TIME/DIV switch to the fastest sweep speed that provides a usable (visible) intensified zone.

Use the DELAY-TIME POSITION dial to move the start of the intensified zone (left-hand edge) to just touch the intersection of the signal and the 10% graticule line (see Fig. 24, Point A). Note the DELAY TIME POSITION dial setting.



Fig. 24. Risetime.

Use the DELAY TIME POSITION dial to move the start of the intensified zone to just touch the intersection of the signal and the 90% graticule line (see Fig. 24, Point B). Note the DELAY TIME POSITION dial setting.

Substitute the DELAY TIME POSITION dial settings in the time difference formula to find the risetime.

EXAMPLE: The A TIME/DIV switch is set to 1 μs. The DELAY TIME POSITION dial setting at point A (Fig. 24) is 2.50. The DELAY TIME POSITION dial setting at point B (Fig. 24) is 7.50. To find the risetime use the formula:

$$\text{Time Difference} = \text{Rise time} = \left[\begin{array}{c}\text{second dial setting} \\ - \\ \text{first dial setting}\end{array}\right] \times \begin{array}{c}\text{delay time} \\ \text{A TIME/DIV} \\ \text{setting}\end{array}$$

Substituting the given values:

Risetime = [7.50 − 2.50] X 1 μs = 5 microseconds.

Mix (Basic 465)

For the MIX mode of operation, the same general procedures can be used. With the first part of the display at a sweep rate set by the A TIME/DIV switch and the second part of the display at a sweep rate set by the B TIME/DIV switch, it is not necessary to switch display modes to ensure location of the correct pulse.

However, inaccuracies are introduced into the measurement by the transition from A to B sweeps. The B DLY'D mode is the most accurate and therefore recommended mode of making differential time measurements.

DM43/DM40 DELAYED OR MIXED SWEEP TIME MEASUREMENTS

Most measurements of time, time-duration, frequency, time difference and risetime are more easily performed using the TIME function of the DM43/DM40 and the delayed-sweep mode of the oscilloscope.

Time Difference Between Repetitive Pulses (DM43/DM40)

Set the HORIZ DISPLAY switch to A INT and the B Trigger SOURCE to STARTS AFTER DELAY. When possible set the B TIME/DIV switch setting 20 or more times faster than the A TIME/DIV switch setting. Position the display as shown in Fig. 25A. Use the DELAY-TIME POSITION knob to move the intensified portion to the first pulse.

Set the HORIZ DISPLAY switch to B DLY'D. Adjust the DELAY-TIME POSITION knob to move the pulse (or rising portion) to some vertical reference (see Fig. 25B). Push the ZERO button on the DM43/DM40 to set the readout to zero.



Fig. 25. Time difference between repetitive pulses.

Turn the DELAY-TIME POSITION knob clockwise to position the second pulse to this same point (if several pulses are displayed, return to the A INT position to locate the correct pulse). The readout display and the TIME lamps indicate the value measured.

EXAMPLE: The A TIME/DIV switch is set to 2 ms and the B TIME/DIV switch is set to .1 ms. Use the DELAY TIME POSITION knob to move point A (see Fig. 25) to the reference line and push the ZERO button on the DM43/DM40. Use the DELAY TIME POSITION knob to move point B to the reference line. The readout is 12 and the ms lamp is lit.

Time Duration (DM43/DM40)

Be sure the VAR TIME/DIV control is set to the calibrated detent. Set the A TIME/DIV switch for a single event and the B TIME/DIV switch 20 times faster (when possible). The most accurate readings are obtained when the B TIME/DIV switch is set to the fastest sweep speed that gives a usable (visible) intensified zone. Position the display to place the time measurement points to the center horizontal line (see Fig. 26).

Use the DELAY-TIME POSITION knob to move the intensified zone (left-hand edge) to just touch the intersection of the signal and the center horizontal line. Push the ZERO button on the DM43/DM50 to set the readout to zero.



Fig. 26. Time duration.

Use the DELAY-TIME POSITION knob to move the intensified zone so its start just touches the second time measurement point.

The readout display and the TIME lamps indicate the value measured.

EXAMPLE: The A TIME/DIV switch is set to 2 ms and the B TIME/DIV switch is set to .1 ms. The ZERO button is pushed at A (see Fig. 26).

The DELAY-TIME POSITION knob sets the start of the intensified portion to B.

The readout is 16.6 and the ms lamp is lit.

Frequency (DM43/DM40)

Time measurements can be used to determine the frequency of a recurrent signal as follows:

1. Measure one cycle of the waveform as described in Time-Duration Measurements.

2. Take the reciprocal of the time duration to determine the frequency.

EXAMPLE: The frequency of the signal shown in Fig. 26, which has a time duration of 16.6 milliseconds is:

$$\text{Frequency} = \frac{1}{\text{time duration}} = \frac{1}{16.6\ \text{ms}} = 60\ \text{hertz}$$

Time Difference Between Two Pulses From Different Sources (DM43/DM40)

Make sure the VAR TIME/DIV control is in the calibrated detent. Set the A Trigger SOURCE switch to CH 1. Connect the reference signal to CH 1 and the comparison signal to CH 2. Connect the signals to the input connectors using probes or cables with equal time delay.

Set the VERT MODE switch to either CHOP or ALT. In general, CHOP is more suitable for low-frequency signals and the ALT position is more suitable for high-frequency signals. Center each of the displays vertically (see Fig. 27A).

Set the HORIZ DISPLAY switch to A INT and the B Trigger SOURCE to STARTS AFTER DELAY. Set the B TIME/DIV switch 20 times faster than the A TIME/DIV switch (when possible). Use the DELAY-TIME POSITION knob to move the channel 1 pulse (or the rising portion) to some vertical reference line (see Fig. 27B). Push the ZERO button on the DM43/DM40 to set the readout to zero.



Fig. 27. Time difference between two pulses from different sources.

Turn the DELAY-TIME POSITION knob clockwise to position the Channel 2 pulse to this same point. The readout display and the TIME lamps indicate the value measured.

EXAMPLE: The A TIME/DIV switch is set to 50 μs and the B TIME/DIV switch is set to 2 μs. Use the DELAY-TIME POSITION knob to position the Channel 1 signal to some vertical reference line (see Fig. 27B). The ZERO button is pushed at A (for Channel 1).

The DELAY-TIME POSITION knob sets the Channel 2 signal to the same vertical reference line (see Fig. 27B).

The readout is 22.5 and the μs lamp is lit.

Risetime (DM43/DM40)

Risetime measurements use the same methods as time-duration measurements, except the measurements are made between the 10% and 90% points of the waveform. Falltime is measured between the 90% and 10% points on the trailing edge of the waveform.

Use a sweep speed setting that displays several cycles or events (if possible) and be sure the VAR TIME/DIV control is in the calibrated detent. Set the VOLTS/DIV switch and the VAR control (or signal amplitude) for exactly a five-division display. Set vertical positioning so the display bottom touches the 0% graticule line and the display top touches the 100% graticule line.

Set the A TIME/DIV switch for a single-event display and horizontally position the display so the 10% point of the waveform intersects the second vertical graticule line (see Fig. 28). Set the B TIME/DIV switch to the fastest setting that provides a usable (visible) intensified zone.



Fig. 28. Risetime.

Use the DELAY-TIME POSITION knob to move the intensified zone start (left-hand edge) to just touch the intersection of the signal and the 10% graticule line (see Fig. 28). Push the ZERO button so the DM43/DM40 sets the readout to zero.

Use the DELAY-TIME POSITION knob to move the intensified zone so its start just touches the 90% graticule line.

The readout is 5.00 and the ms lamp is lit.

Mix (DM43/DM40)

For the MIX mode of operation, the same general procedures can be used. With the first part of the display at a sweep rate set by the A TIME/DIV switch, and the second part of the display at a sweep rate set by the B TIME/DIV switch, it is not necessary to switch display modes to ensure location of the correct pulse.

However, inaccuracies are introduced into the measurement by the transition from A to B sweeps. The B DLY'D mode is the most accurate and therefore the recommended mode of making differential time measurements.

DELAYED OR MIXED SWEEP MAGNIFICATION

The delayed sweep features of the 465 can be used to provide higher apparent magnification than is provided by the X10 MAG switch. The sweep rate of the delayed sweep (B sweep) is not actually increased; the apparent magnification is the result of delaying the B sweep an amount of time selected by the A TIME/DIV switch and the DELAY-TIME POSITION control before the display is presented at the sweep rate selected by the B TIME/DIV switch. The following method uses the STARTS AFTER DELAY position of the B Trigger SOURCE switch to allow the delayed portion to be positioned with the DELAY-TIME POSITION control. If there is too much jitter in the delayed display, use the Triggered B Sweep mode of operation.

Magnified Sweep Starts After Delay

1. Connect the signal to either input connector. Set the VERT MODE switch to display the channel used.

2. Set the VOLTS/DIV switch to produce a display about four divisions in amplitude.

3. Set the A TIME/DIV switch to a sweep rate which displays the complete waveform.

4. Set the HORIZ DISPLAY switch to A INT and the B Trigger SOURCE switch to STARTS AFTER DELAY.

5. Position the start of the intensified zone with the DELAY-TIME POSITION control to the part of the display to be magnified.

6. Set the B TIME/DIV switch to a setting which intensifies the full portion to be magnified. The start of the intensified zone remains as positioned above (see Fig. 29A).

7. Set the HORIZ DISPLAY switch to B DLY'D to magnify the portion of A sweep that is intensified (see Fig. 29B).

8. Time measurements can be made from the display in the conventional manner. The sweep rate is determined by the setting of the B TIME/DIV switch.

9. The apparent sweep magnification can be calculated by dividing the A TIME/DIV switch setting by the B TIME/DIV switch setting.



Fig. 29. Delayed sweep magnification.

EXAMPLE: The apparent magnification of a display with an A TIME/DIV switch setting of .1 ms and a B TIME/DIV switch setting of 1 μs is:

$$\text{Apparent Magnification (Delayed Sweep)} = \frac{\text{A TIME/DIV setting}}{\text{B TIME/DIV setting}}$$

Substituting the given values:

$$\text{Apparent Magnification} = \frac{1 \times 10^{-4}}{1 \times 10^{-6}}$$

The apparent magnification is 100 times.

Triggered Delayed Sweep Magnification

The delayed sweep magnification method just described may produce too much jitter at high apparent magnification ranges. Operating the B Sweep in a triggered mode provides a more stable display since the delayed display is triggered at the same point each time.

1. Set up the display as given in steps 1 through 6 under "Magnified Sweep Starts After Delay".

2. Set the B Trigger SOURCE switch to the same position as the A Trigger SOURCE switch.

3. Adjust the B LEVEL control so the intensified zone on the trace is stable. (If an intensified zone cannot be obtained, see step 4.)

4. Inability to intensify the desired portion indicates that the signal does not meet the triggering requirements. If the condition cannot be remedied with the B Triggering controls or by increasing the display amplitude (lower VOLTS/DIV setting), trigger B Sweep externally.

5. When the correct portion is intensified, set the HORIZ DISPLAY switch to B DLY'D. Slight readjustment of the B LEVEL control may be necessary for a stable display.

6. Measurements are made and magnification factors are calculated as in the STARTS AFTER DELAY mode previously given.

PHASE DIFFERENCE MEASUREMENTS

Use either the CHOP or ALT mode. Set the A Trigger SOURCE switch to CH 1. The reference signal should precede the comparison signal in time. Use coaxial cables or probes that have equal time delay to connect the signals to the input connectors.

If the signals are of opposite polarity, set the INVERT pushbutton to invert the Channel 2 display (signals may be of opposite polarity due to 180° phase difference; if so, take this into account in the final calculation). Set the CH 1 and CH 2 VOLTS/DIV switches and the CH 1 and CH 2 VAR controls so the displays are equal in amplitude.

Set the TIME/DIV switch for about a one-cycle waveform. Position the display and turn the A VAR TIME/DIV control for 1 reference signal cycle in exactly eight divisions (see Fig. 30). Each division of the graticule represents 45° of the cycle (360° ÷ 8 divisions = 45°/division). The sweep rate can be stated in terms of degrees as 45°/division.

Measure the horizontal difference between corresponding points on the waveforms and multiply the distance measured (in divisions) by 45°/division (sweep rate) to obtain the amount of phase difference.



Fig. 30. Phase difference.

EXAMPLE: The horizontal difference is 0.6 division with a sweep rate of 45°/division as shown in Fig. 30.

Using the formula:

$$\text{Phase Difference} = \text{horizontal difference divisions} \times \text{sweep (degrees/div)}$$

Substituting the given values:

Phase Difference = 0.6 X 45° = 27°.

HIGH RESOLUTION PHASE DIFFERENCE MEASUREMENTS

Make more accurate phase measurements by increasing the sweep rate (without changing the A VAR TIME/DIV control) by using the X10 MAG mode. Delayed sweep magnification may also be used (see Fig. 31).

EXAMPLE: If the sweep rate were increased 10 times with the magnifier, the magnified sweep rate would be 45° ÷ 10 = 4.5°/division. Fig. 31 shows the same signals as used in Fig. 30 but with the X10 MAG switch set to X10. With a horizontal difference of 6 divisions, the phase difference is:

$$\text{Phase Difference} = \text{horizontal difference (divisions)} \times \text{magnified sweep rate (degrees/div)}$$

Substituting the given values:

Phase Difference = 6 X 4.5° = 27°.



Fig. 31. High resolution phase difference.

PULSE JITTER MEASUREMENTS

Be sure the VAR TIME/DIV switch is in the calibrated detent. Set the B TIME/DIV switch to intensify the full rising portion of the pulse. Set the HORIZ DISPLAY switch to B DLY'D.

Pulse jitter is shown by horizontal movement of the pulse and includes the inherent jitter of the Delayed Sweep (see Fig. 32). Multiply the distance by the B TIME/DIV switch setting to obtain pulse jitter time.



Fig. 32. Pulse jitter.

OPERATOR'S SPECIFICATIONS

Refer to the service manual for complete specifications. Specifications given are for an operating range of 0° C to +40° C unless otherwise stated.

VERTICAL

Deflection Factor Accuracy: Within 3% in the calibrated position.

Frequency Response: Dc to at least 100 MHz for CH 1 and CH 2. Ac-coupled, low-frequency response is 10 Hz or less. Use of a 10X probe extends frequency response to 1 Hz.

Risetime: 3.5 nanoseconds or less (calculated from 0.35 ÷ bandwidth in MHz).

Maximum Input Voltage: Dc coupled, 250 V (dc + peak ac) or 500 V p-p ac at 1 kHz or less,; ac coupled, 500 V (dc + peak ac) or 500 V p-p ac 1 kHz or less.

Positive-Going Step Aberrations: Less than +3%, −3%, not to exceed 3% peak-to-peak, excluding the ADD mode.

Common-Mode Rejection Ratio (ADD Mode with CH 2 Inverted): At least 10:1 at 20 MHz for common mode signals of 6 divisions or less with GAIN adjusted for best CMRR at 50 kHz.

Input Gate Current: 0.5 nA or less (0.1 div at 5 mV/DIV), from −15° C to +30° C.

Channel Isolation: At least 100:1 at 25 MHz.

CHOP Mode Repetition Rate: Approximately 250 kHz.

Cascaded Operation (CH 1 OUT into CH 2) sensitivity: At least 1 mV/DIV. Bandwidth is dc to at least 50 MHz, with CH 1 OUT connected to CH 2 Input, ac coupled, using a 50 Ω, 42-inch BNC cable, terminated in 50 Ω at CH 2 Input.

TRIGGERING

Sensitivity

Ac Coupled Signal: 0.3 div internal or 50 mV external, from 30 Hz to 25 MHz; increasing to 1.5 div. internal or 150 mV external at 100 MHz.

LF REJ Coupled Signal: 0.5 div internal or 100 mV external, from 50 kHz to 25 MHz; increasing to 1.5 div internal or 300 mV external at 100 MHz. Attenuates signals below about 50 kHz.

HF REJ Coupled Signal: 0.5 div internal or 50 mV external, from 30 Hz to 50 kHz. Attenuates signals below about 60 Hz and above about 50 kHz.

DC Coupled Signal: 0.3 div internal or 50 mV external, from dc to 25 MHz; increasing to 1.5 div internal or 150 mV external at 100 MHz.

EXT ÷ 10 Signal: Amplitude requirements are multiplied by 10.

External Trigger Input

Maximum Input Voltage: 250 V (dc + peak ac) or 250 V p-p ac (1 kHz or less).

Level Control Range In Ext

At least + and −2 V, 4 V p-p; EXT ÷ 10 is at least + and −20 V, 40 V p-p.

Trigger View

Deflection Factor: About 50 mV/div in EXT and about 500 mV/div in EXT ÷ 10.

Risetime: 5 ns or less.

Trigger Centering Point: Within 1.0 division of screen center.

DIFFERENTIAL TIME MEASUREMENT (BASIC 465)

Accuracy for Measurements Greater than One Major Dial Division: Within ±1% from +15° C to +35° C. Within ±2.5% from −15° C to +55° C.

Accuracy for Measurements Less than One Major Dial Division: Within ±0.01 major dial divisions from +15° C to +35° C. Within ±0.03 major dial divisions from −15° C to +55° C.

DIFFERENTIAL TIME MEASUREMENT (DM43/DM40)

Accuracy: +15° C to +35° C; within 0.75% of reading, ±1 count. −15° C to +55° C; within 1.25% of reading, ±1 count.

Zero Reference: Intial Reset, 0000, ±1 digit.

HORIZONTAL

Sweep Rate Accuracy: Within 2%, unmagnified, and 3% magnified, from +20° C to +30° C for A and B Sweeps.

Mixed Sweep Accuracy: Within 2% plus the measured A Sweep inaccuracy, when viewing the A portion only. B Sweep portion remains the same as above.

Trigger Holdoff Variable: Increases A Sweep holdoff time by at least a factor of 10.

Delay Time Jitter: 1 part, or less, in 50,000 (0.002%) of 10 times the A TIME/DIV switch setting.

Calibrated Delay Time (VAR control to CAL): Continuous from 0.2 μs to at least 5 seconds after the start of A (delaying) sweep.

X-Y

X-Axis Sensitivity (X10 MAG turned off): Same as the vertical system.

X-Axis Bandwidth: Dc to at least 4 MHz, with a 10 division reference signal.

Phase Difference Between X- and Y-Axis Amplifiers: Within 3° from dc to 50 kHz.

Deflection Accuracy: Within 4%.

CALIBRATOR

Output Voltage: 0.3 V within 1% and within 0.3% from +20° C to +30° C.

Output Current: 30 mA within 2% from +20° C to +30° C.

Repetition Rate: About 1 kHz.

Z AXIS INPUT

Sensitivity: Noticeable intensity modulation, at normal intensity settings, by a 5 V p-p signal. A positive-going signal decreases intensity.

Frequency Range (Usable): Dc to 50 MHz.

Maximum Input Voltage: 100 V (dc plus peak ac) or 100 V p-p ac at 1 kHz or less.

OUTPUTS

CH 1 Out

Output Voltage: At least 50 mV/div into 1 M Ω; to at least 25 mV/div into 50 ohms.

Bandwidth: Dc to at least 50 MHz into 50 ohms.

DC Level: About 0 volts.

A+ and +B Gates

Output Voltage: About 5.5 V of positive-going pulse.

Output Resistance: About 500 ohms.

RESISTANCE DM43/DM40

Maximum Safe Input Voltage: 120 V rms between + and COM inputs.

Resistance Accuracy: 20 kΩ, 200 kΩ, and 2 MΩ ranges—within 0.3%, ±1 count; 200 Ω, 2 kΩ, and 20 MΩ ranges—within 0.75%, ±1 count.

Temperature Dependence: 20 kΩ, 200 kΩ and 2 MΩ ranges—250 ppm/° C; 200 Ω, 2 kΩ and 20 MΩ ranges—350 ppm/° C.

Resolution: 0.1 Ω.

Recycle Time: At least 3 measurements/second.

Response Time: All ranges within 1 second except 20 MΩ (within 5 seconds).

TIME DM43/DM40

See DIFFERENTIAL TIME MEASUREMENT (DM43/DM40).

TEMPERATURE—DM43 ONLY, WITH PROBE

Range: −55° C to +150° C in 1 range.

Accuracy: Original Probe — Within 2° C, −55° C to +125° C. Within 3° C, +125° C to +150° C.

Replacement Probe: Accuracy with equal original probe accuracy after DM43 is compensated.

Maximum Safe Voltage on Measurement Surfaces: ±400 V (dc + peak ac) above chassis ground.

Temperature (Storage and Operating): Probe Body and Cable, −55° C to +105° C. Probe Sensor Tip, −55° C to +150° C.

DC VOLTAGE DM43/DM40

Maximum Safe Input Voltage: 200 mV and 2 V ranges; ±500 V (dc + peak ac) between + and COM inputs. 20 V, 200 V, and 1.2 kV ranges; ±1200 V (dc + peak ac) between + and chassis.

Common Floating Voltage: ±500 V (dc + peak ac) to chassis.

DC Voltage Accuracy: Within 0.1% of reading, ±1 count.

Temperature Dependence: 44 ppm/° C.

Resolution: 100 μV.

Recycle Time: At least 3 measurements/second.

Response Time: Within 0.5 second.

Normal/Common Mode Rejection Ratio:

Normal Mode—At least 60 dB at 50 Hz and 60 Hz.

Common Mode—At least 100 dB at dc; 80 dB at 50 Hz and 60 Hz.

Input Impedance: 200 mV and 2 V ranges, about 1000 MΩ. 20 V, 200 V, and 1.2 kV ranges, about 10 MΩ.

AC POWER SOURCE

115 V		230 V	
Low	99 V to 121 V	Low	198 V to 242 V
Medium	104 V to 126 V	Medium	207 V to 253 V
High	108 V to 132 V	High	216 V to 264 V

Line Frequency: From 48 Hz to 440 Hz.

Maximum Power Consumption: 100 watts at 115 V, 60 Hz, medium range,

ENVIRONMENTAL

Operating Temperature: −15°C to +55°C.

Operating Altitude: To 15,000 feet. Maximum operating temperature decreases 1°C/1,000 feet, above 5,000 feet. Storage temperature is to 50,000 feet.

Humidity (Operating and Storage): 5 cycles (120 hours) referenced to MIL-E-16400F.

ACCESSORIES

STANDARD ACCESSORIES INCLUDED

2	P6065A Probe Packages	010-6065-13
1	Accessory Pouch, Snap (w/o DM)	016-0535-02
1	Accessory Pouch, Snap (with DM)	016-0594-00
1	Accessory Pouch, Zipper	016-0537-00
1	Operator's Manual	070-1738-01
1	Service Manual (465)	070-1861-00
2	Fuses, 1.5 A 3AG fast-blow	159-0016-00
2	Fuses, 0.75 A 3AG fast-blow	159-0042-00
1	Filter, Blue Plastic (installed)	337-1674-00
1	CRT Filter, Clear Plastic	337-1674-01
1	Adapter, Ground Wire	134-0016-01
1	Pair, Test Leads (with DM)	003-0120-00
1	P6430 Probe (DM43 Temp.)	010-6430-00
1	Service Manual (DM43/DM40)	070-1779-00

OPTIONAL ACCESSORIES

C30A-P Option 1 Compact Camera – f1.9 lens, 0.8 magnification, Polaroid Land Pack Film Back for 3000-speed film (includes Adapter Frame/Corrector Lens 016-0301-00) – Order C30A-P Option 1.

Camera Adapter and Corrector Lens—Adapts C-30A to 464, 465, 466, or 475—Order 016-0301-00.

Protective Cover—Waterproof, blue vinyl – Order 016-0554-00.

Folding Polarized Viewing Hood—Order 016-0180-00.

Folding Viewing Hood, light-shielding – Order 016-0592-00.

Folding Viewing Hood, light-occluding – Order 016-0566-00.

Mesh Filter—Improves contrast and EMI filter – Order 378-0726-01.

SCOPE-MOBILE Cart—Occupies less than 18 inches aisle space, has storage area in base – Order 200.

OPTIONS

Your instrument may contain the following options:

OPTION 4

The instrument is modified to meet certain specifications on radiated interference requirements. There is no change in operating instructions.

OPTION 7

At the time of this writing, instruments having the DM43 or DM40 do not have the Option 7 available.

Option 7 permits operation on 12 or 24 Vdc with no performance deterioration. Circuitry is provided to protect against damage due to connection of 24 V when in the 12 V mode of operation. The 24- volt external input permits use with conventional dc power (marine and aircraft).